

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071621\
 Data File : VN067778.D
 Acq On : 16 Jul 2021 17:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 17 04:59:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	472524	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	860674	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	812505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	322544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	329601	46.154	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.300%
35) Dibromofluoromethane	8.024	113	251216	48.311	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.620%
50) Toluene-d8	10.443	98	1055566	48.496	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.000%
62) 4-Bromofluorobenzene	12.734	95	363228	47.241	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.480%
Target Compounds						
					Qvalue	
95) Naphthalene	15.509	128	24663	4.480	ug/l	96
96) 1,2,3-Trichlorobenzene	15.703	180	4931	3.539	ug/l	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071621\
Data File : VN067778.D
Acq On : 16 Jul 2021 17:22
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 17 04:59:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

