

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
 Data File : VN082931.D
 Acq On : 16 Jul 2024 13:02
 Operator : JC\MD
 Sample : VN0716WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0716WBL01

Quant Time: Jul 17 00:53:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	75723	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	144953	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	145483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	59445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	61448	57.889	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.780%
35) Dibromofluoromethane	8.171	113	49691	53.424	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.840%
50) Toluene-d8	10.571	98	176587	51.402	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.800%
62) 4-Bromofluorobenzene	12.853	95	54837	44.437	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	88.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
Data File : VN082931.D
Acq On : 16 Jul 2024 13:02
Operator : JC\MD
Sample : VN0716WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0716WBL01

Quant Time: Jul 17 00:53:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

