

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
 Data File : VN082938.D
 Acq On : 16 Jul 2024 16:20
 Operator : JC\MD
 Sample : P3256-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-TB-20240712

Quant Time: Jul 17 00:57:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.194	168	88673	50.000	ug/l	-0.03
34) 1,4-Difluorobenzene	9.083	114	171973	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.865	117	163004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	62031	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.553	65	62141	49.992	ug/l	-0.03
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.980%
35) Dibromofluoromethane	8.136	113	52061	47.178	ug/l	-0.03
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	10.559	98	178241	43.732	ug/l	-0.01
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.460%
62) 4-Bromofluorobenzene	12.847	95	53009	36.207	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	72.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
Data File : VN082938.D
Acq On : 16 Jul 2024 16:20
Operator : JC\MD
Sample : P3256-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-TB-20240712

Quant Time: Jul 17 00:57:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

