

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
 Data File : VN082942.D
 Acq On : 16 Jul 2024 18:00
 Operator : JC\MD
 Sample : P3246-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CTW2-BL-02

Quant Time: Jul 17 00:59:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	80218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	171854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	184967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72395	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	63120	56.132	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.260%
35) Dibromofluoromethane	8.171	113	52004	47.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	10.571	98	203593	49.987	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.980%
62) 4-Bromofluorobenzene	12.853	95	65734	44.930	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	89.860%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	27099	68.448	ug/l	98
20) Methylene Chloride	5.271	84	7624	7.974	ug/l #	80
43) Isopropyl Acetate	8.694	43	116416	39.248	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
Data File : VN082942.D
Acq On : 16 Jul 2024 18:00
Operator : JC\MD
Sample : P3246-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CTW2-BL-02

Quant Time: Jul 17 00:59:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

