

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN071624\  
 Data File : VN082943.D  
 Acq On : 16 Jul 2024 18:25  
 Operator : JC\MD  
 Sample : P3246-10  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 WC-CTW2-BL-02-G

Quant Time: Jul 17 00:59:30 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070824W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 08 13:46:04 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min)     |
|-----------------------------|--------|-------|----------|----------|--------|--------------|
| -----                       |        |       |          |          |        |              |
| Internal Standards          |        |       |          |          |        |              |
| 1) Pentafluorobenzene       | 8.229  | 168   | 86230    | 50.000   | ug/l   | 0.00         |
| 34) 1,4-Difluorobenzene     | 9.106  | 114   | 181848   | 50.000   | ug/l   | 0.00         |
| 63) Chlorobenzene-d5        | 11.870 | 117   | 194595   | 50.000   | ug/l   | 0.00         |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 72468    | 50.000   | ug/l   | 0.00         |
| System Monitoring Compounds |        |       |          |          |        |              |
| 33) 1,2-Dichloroethane-d4   | 8.582  | 65    | 64939    | 53.723   | ug/l   | 0.00         |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 107.440%     |
| 35) Dibromofluoromethane    | 8.171  | 113   | 54695    | 46.873   | ug/l   | 0.00         |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 93.740%      |
| 50) Toluene-d8              | 10.570 | 98    | 205370   | 47.652   | ug/l   | 0.00         |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 95.300%      |
| 62) 4-Bromofluorobenzene    | 12.853 | 95    | 64388    | 41.591   | ug/l   | 0.00         |
| Spiked Amount               | 50.000 | Range | 64 - 133 | Recovery | =      | 83.180%      |
| Target Compounds            |        |       |          |          |        |              |
| 16) Acetone                 | 4.436  | 43    | 21252    | 49.937   | ug/l   | Qvalue<br>99 |
| 20) Methylene Chloride      | 5.288  | 84    | 6086     | 5.922    | ug/l # | 82           |
| 25) 2-Butanone              | 7.500  | 43    | 3351     | 5.215    | ug/l   | 94           |
| 43) Isopropyl Acetate       | 8.694  | 43    | 114288   | 36.186   | ug/l # | 85           |
| -----                       |        |       |          |          |        |              |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN071624\  
Data File : VN082943.D  
Acq On : 16 Jul 2024 18:25  
Operator : JC\MD  
Sample : P3246-10  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
WC-CTW2-BL-02-G

Quant Time: Jul 17 00:59:30 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070824W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 08 13:46:04 2024  
Response via : Initial Calibration

