

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
 Data File : VN082945.D
 Acq On : 16 Jul 2024 19:15
 Operator : JC\MD
 Sample : P3246-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CTW2-BL-03-G

Quant Time: Jul 17 01:00:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	86221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	194825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	75794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	68385	56.580	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.160%
35) Dibromofluoromethane	8.165	113	55136	46.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	10.571	98	212780	48.969	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.940%
62) 4-Bromofluorobenzene	12.853	95	68242	43.722	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.440%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	19694	46.281	ug/l	91
20) Methylene Chloride	5.283	84	5887	5.729	ug/l #	74
43) Isopropyl Acetate	8.694	43	104770	32.616	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071624\
Data File : VN082945.D
Acq On : 16 Jul 2024 19:15
Operator : JC\MD
Sample : P3246-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CTW2-BL-03-G

Quant Time: Jul 17 01:00:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

