

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071718\
 Data File : VN049852.D
 Acq On : 17 Jul 2018 17:52
 Operator : MD\SY
 Sample : PB111173TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111173TB

Quant Time: Jul 18 11:37:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 17 13:20:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	632726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	971244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	849469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	322395	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	427288	50.42	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.84%	
35) Dibromofluoromethane	7.59	113	373484	48.48	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.96%	
50) Toluene-d8	10.09	98	1355363	47.92	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	12.40	95	386841	40.84	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	81.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
29) Tetrahydrofuran	7.22	42	2351	1.06	ug/l #	75
43) Isopropyl Acetate	8.17	43	256266	19.65	ug/l	97
59) 2-Hexanone	10.77	43	25359	4.82	ug/l #	57
95) Naphthalene	15.14	128	10391	3.53	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071718\
Data File : VN049852.D
Acq On : 17 Jul 2018 17:52
Operator : MD\SY
Sample : PB111173TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111173TB

Quant Time: Jul 18 11:37:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 17 13:20:18 2018
Response via : Initial Calibration

