

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071718\
 Data File : VN049853.D
 Acq On : 17 Jul 2018 18:18
 Operator : MD\SY
 Sample : PB111173ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111173ZHE#01

Quant Time: Jul 18 11:37:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 17 13:20:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	608966	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	950002	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	836471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	327977	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	415148	50.89	ug/l	0.00
Spiked Amount			Recovery	=	101.78%	
35) Dibromofluoromethane	7.59	113	365769	48.54	ug/l	0.00
Spiked Amount			Recovery	=	97.08%	
50) Toluene-d8	10.09	98	1312451	47.44	ug/l	0.00
Spiked Amount			Recovery	=	94.88%	
62) 4-Bromofluorobenzene	12.40	95	376287	40.61	ug/l	0.00
Spiked Amount			Recovery	=	81.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
43) Isopropyl Acetate	8.17	43	253872	19.90	ug/l	96
59) 2-Hexanone	10.77	43	25543	4.97	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071718\
Data File : VN049853.D
Acq On : 17 Jul 2018 18:18
Operator : MD\SY
Sample : PB111173ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111173ZHE#01

Quant Time: Jul 18 11:37:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 17 13:20:18 2018
Response via : Initial Calibration

