

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071718\
 Data File : VN049858.D
 Acq On : 17 Jul 2018 20:28
 Operator : MD\SY
 Sample : J3994-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGEBLANK-WATERREF-4

Quant Time: Jul 18 07:19:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 17 13:20:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	570522	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	919753	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	763330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	292740	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	405697	53.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.18%	
35) Dibromofluoromethane	7.59	113	355811	48.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.54%	
50) Toluene-d8	10.09	98	1234751	46.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.20%	
62) 4-Bromofluorobenzene	12.40	95	341590	38.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071718\
Data File : VN049858.D
Acq On : 17 Jul 2018 20:28
Operator : MD\SY
Sample : J3994-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGEBLANK-WATERREF-4

Quant Time: Jul 18 07:19:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 17 13:20:18 2018
Response via : Initial Calibration

