

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
 Data File : VN056792.D
 Acq On : 17 Jul 2019 13:24
 Operator : JC/SP
 Sample : K3847-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW-7

Quant Time: Jul 18 01:49:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	250593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	450852	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	465835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	160022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	149544	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
35) Dibromofluoromethane	7.59	113	132304	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
50) Toluene-d8	10.09	98	566072	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
62) 4-Bromofluorobenzene	12.40	95	200016	54.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.46%	
Target Compounds						
16) Acetone	3.83	43	6679	5.229	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071719\
Data File : VN056792.D
Acq On : 17 Jul 2019 13:24
Operator : JC/SP
Sample : K3847-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW-7

Quant Time: Jul 18 01:49:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

