

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
 Data File : VN056795.D
 Acq On : 17 Jul 2019 14:34
 Operator : JC/SP
 Sample : K3841-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS027-190715

Quant Time: Jul 18 08:22:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	254985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	451289	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	473349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	170553	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	147012	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
35) Dibromofluoromethane	7.59	113	135953	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
50) Toluene-d8	10.09	98	564066	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	12.40	95	204449	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.83	96	103124	35.371	ug/l	91
44) Trichloroethene	8.84	130	16305	4.692	ug/l	99
64) Tetrachloroethene	10.63	164	12081	2.958	ug/l	97
95) Naphthalene	15.13	128	5423	5.065	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071719\
Data File : VN056795.D
Acq On : 17 Jul 2019 14:34
Operator : JC/SP
Sample : K3841-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS027-190715

Quant Time: Jul 18 08:22:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

