

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
 Data File : VN056796.D
 Acq On : 17 Jul 2019 14:57
 Operator : JC/SP
 Sample : K3841-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS037-190715

Quant Time: Jul 18 08:23:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	267354	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	478601	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	505619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182370	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	156921	53.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.12%	
35) Dibromofluoromethane	7.59	113	141388	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
50) Toluene-d8	10.09	98	606363	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
62) 4-Bromofluorobenzene	12.40	95	219219	56.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.02%	
Target Compounds						
16) Acetone	3.82	43	12489	9.164	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071719\
Data File : VN056796.D
Acq On : 17 Jul 2019 14:57
Operator : JC/SP
Sample : K3841-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS037-190715

Quant Time: Jul 18 08:23:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

