

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
 Data File : VN056799.D
 Acq On : 17 Jul 2019 16:07
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 18 09:12:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.469	0.415	11.5	84	0.00
3 P	Chloromethane	0.631	0.550	12.8	84	0.00
4 C	Vinyl Chloride	0.608	0.548	9.9#	86	0.00
5 T	Bromomethane	0.357	0.305	14.6	80	0.00
6 T	Chloroethane	0.354	0.315	11.0	86	0.00
7 T	Trichlorofluoromethane	0.780	0.727	6.8	89	0.00
8 T	Diethyl Ether	0.304	0.298	2.0	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.469	0.445	5.1	89	0.00
10 T	Methyl Iodide	0.627	0.524	16.4	74	0.00
11 T	Tert butyl alcohol	0.092	0.095	-3.3	94	0.00
12 CM	1,1-Dichloroethene	0.470	0.435	7.4#	87	0.00
13 T	Acrolein	0.033	0.024	27.3#	77	0.00
14 T	Allyl chloride	0.813	0.780	4.1	88	0.00
15 T	Acrylonitrile	0.266	0.272	-2.3	92	0.00
16 T	Acetone	0.255	0.233	8.6	85	0.00
17 T	Carbon Disulfide	1.242	1.096	11.8	83	0.00
18 T	Methyl Acetate	0.684	0.571	16.5	92	0.00
19 T	Methyl tert-butyl Ether	1.498	1.453	3.0	90	0.00
20 T	Methylene Chloride	0.557	0.528	5.2	89	0.00
21 T	trans-1,2-Dichloroethene	0.497	0.476	4.2	88	0.00
22 T	Diisopropyl ether	1.604	1.588	1.0	89	0.00
23 T	Vinyl Acetate	1.300	1.348	-3.7	90	0.00
24 P	1,1-Dichloroethane	0.930	0.903	2.9	90	0.00
25 T	2-Butanone	0.373	0.369	1.1	90	0.00
26 T	2,2-Dichloropropane	0.671	0.654	2.5	88	0.00
27 T	cis-1,2-Dichloroethene	0.572	0.566	1.0	90	0.00
28 T	Bromochloromethane	0.437	0.440	-0.7	91	0.00
29 T	Tetrahydrofuran	0.243	0.247	-1.6	91	0.00
30 C	Chloroform	0.931	0.893	4.1#	91	0.00
31 T	Cyclohexane	0.884	0.833	5.8	87	0.00
32 T	1,1,1-Trichloroethane	0.736	0.728	1.1	89	0.00
33 S	1,2-Dichloroethane-d4	0.553	0.544	1.6	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.311	0.315	-1.3	91	0.00
36 T	1,1-Dichloropropene	0.463	0.436	5.8	89	0.00
37 T	Ethyl Acetate	0.456	0.479	-5.0	92	0.00
38 T	Carbon Tetrachloride	0.434	0.414	4.6	88	0.00
39 T	Methylcyclohexane	0.571	0.546	4.4	88	0.00
40 TM	Benzene	1.418	1.360	4.1	89	0.00
41 T	Methacrylonitrile	0.200	0.214	-7.0	96	0.00
42 TM	1,2-Dichloroethane	0.448	0.435	2.9	90	0.00
43 T	Isopropyl Acetate	0.717	0.731	-2.0	91	0.00
44 TM	Trichloroethene	0.385	0.367	4.7	90	0.00
45 C	1,2-Dichloropropane	0.377	0.369	2.1#	90	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
 Data File : VN056799.D
 Acq On : 17 Jul 2019 16:07
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 18 09:12:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.233	0.232	0.4	90	0.00
47 T	Bromodichloromethane	0.444	0.443	0.2	90	0.00
48 T	Methyl methacrylate	0.347	0.354	-2.0	90	0.00
49 T	1,4-Dioxane	0.007	0.008	-14.3	100	0.00
50 S	Toluene-d8	1.202	1.204	-0.2	90	0.00
51 T	4-Methyl-2-Pentanone	0.464	0.474	-2.2	92	0.00
52 CM	Toluene	0.842	0.843	-0.1#	90	0.00
53 T	t-1,3-Dichloropropene	0.463	0.479	-3.5	90	0.00
54 T	cis-1,3-Dichloropropene	0.537	0.546	-1.7	90	0.00
55 T	1,1,2-Trichloroethane	0.353	0.343	2.8	91	0.00
56 T	Ethyl methacrylate	0.508	0.539	-6.1	91	0.00
57 T	1,3-Dichloropropane	0.577	0.576	0.2	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.244	0.249	-2.0	86	0.00
59 T	2-Hexanone	0.333	0.349	-4.8	90	0.00
60 T	Dibromochloromethane	0.347	0.355	-2.3	90	0.00
61 T	1,2-Dibromoethane	0.351	0.359	-2.3	91	0.00
62 S	4-Bromofluorobenzene	0.405	0.415	-2.5	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.431	0.391	9.3	89	0.00
65 PM	Chlorobenzene	1.022	0.992	2.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.370	1.6	90	0.00
67 C	Ethyl Benzene	1.797	1.757	2.2#	90	0.00
68 T	m/p-Xylenes	0.671	0.666	0.7	90	0.00
69 T	o-Xylene	0.660	0.645	2.3	89	0.00
70 T	Styrene	1.067	1.102	-3.3	90	0.00
71 P	Bromoform	0.286	0.296	-3.5	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	4.032	3.713	7.9	91	0.00
74 T	N-amyl acetate	1.501	1.444	3.8	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.356	1.194	11.9	92	0.00
76 T	1,2,3-Trichloropropane	1.094	1.072	2.0	94	0.00
77 T	Bromobenzene	1.054	0.978	7.2	91	0.00
78 T	n-propylbenzene	4.476	4.127	7.8	89	0.00
79 T	2-Chlorotoluene	2.769	2.485	10.3	91	0.00
80 T	1,3,5-Trimethylbenzene	3.417	3.136	8.2	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.354	0.357	-0.8	91	0.00
82 T	4-Chlorotoluene	2.627	2.430	7.5	89	0.00
83 T	tert-Butylbenzene	2.974	2.664	10.4	91	0.00
84 T	1,2,4-Trimethylbenzene	3.338	3.149	5.7	91	0.00
85 T	sec-Butylbenzene	3.797	3.506	7.7	90	0.00
86 T	p-Isopropyltoluene	3.440	3.211	6.7	90	0.00
87 T	1,3-Dichlorobenzene	1.728	1.599	7.5	89	0.00
88 T	1,4-Dichlorobenzene	1.681	1.571	6.5	89	0.00
89 T	n-Butylbenzene	2.771	2.656	4.2	88	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071719\
Data File : VN056799.D
Acq On : 17 Jul 2019 16:07
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 18 09:12:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.556	0.525	5.6	91	0.00
91 T	1,2-Dichlorobenzene	1.767	1.631	7.7	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.214	0.9	93	0.00
93 T	1,2,4-Trichlorobenzene	0.810	0.847	-4.6	86	0.00
94 T	Hexachlorobutadiene	0.629	0.560	11.0	94	0.00
95 T	Naphthalene	2.171	2.244	-3.4	85	0.00
96 T	1,2,3-Trichlorobenzene	0.836	0.855	-2.3	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6