

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
 Data File : VN078534.D
 Acq On : 17 Jul 2023 11:08
 Operator : JC\MD
 Sample : VN0717MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0717MBL01

Quant Time: Jul 18 04:19:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	411336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	735700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	681994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	244772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	272957	49.499	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.000%
35) Dibromofluoromethane	8.177	113	278635	55.792	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.580%
50) Toluene-d8	10.576	98	854973	45.317	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.640%
62) 4-Bromofluorobenzene	12.859	95	299603	51.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
Data File : VN078534.D
Acq On : 17 Jul 2023 11:08
Operator : JC\MD
Sample : VN0717MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0717MBL01

Quant Time: Jul 18 04:19:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

