

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
 Data File : VN078538.D
 Acq On : 17 Jul 2023 13:11
 Operator : JC\MD
 Sample : 03637-01DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A508DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/18/2023
 Supervised By : Mahesh Dadoda 07/18/2023

Quant Time: Jul 18 04:21:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	412813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	735814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	693825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	249180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	271417	49.044	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.080%		
35) Dibromofluoromethane	8.171	113	277150	55.486	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	110.980%		
50) Toluene-d8	10.571	98	858738	45.510	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.020%		
62) 4-Bromofluorobenzene	12.853	95	302827	52.102	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.200%		
Target Compounds					Qvalue	
16) Acetone	4.442	43	212395	123.206	ug/l	97
25) 2-Butanone	7.506	43	10799	3.979	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
Data File : VN078538.D
Acq On : 17 Jul 2023 13:11
Operator : JC\MD
Sample : 03637-01DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A508DL

Quant Time: Jul 18 04:21:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 07/18/2023
Supervised By :Mahesh Dadoda 07/18/2023

