

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
 Data File : VN078542.D
 Acq On : 17 Jul 2023 14:49
 Operator : JC\MD
 Sample : 03632-06ME
 Misc : 12.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-09-(3-5)ME

Quant Time: Jul 18 04:24:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	416338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	734876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	745600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	326825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	267923	48.002	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.000%
35) Dibromofluoromethane	8.171	113	276623	55.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.900%
50) Toluene-d8	10.571	98	879493	46.669	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.340%
62) 4-Bromofluorobenzene	12.853	95	370279	63.789	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	127.580%#
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.606	83	11826	1.799	ug/l #	72
67) Ethyl Benzene	11.970	91	40725	1.446	ug/l	90
78) n-propylbenzene	13.041	91	143794	4.378	ug/l	98
89) n-Butylbenzene	14.064	91	46042	3.078	ug/l #	85
95) Naphthalene	15.647	128	126362	9.871	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
Data File : VN078542.D
Acq On : 17 Jul 2023 14:49
Operator : JC\MD
Sample : 03632-06ME
Misc : 12.08g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-09-(3-5)ME

Quant Time: Jul 18 04:24:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

