

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
 Data File : VN078548.D
 Acq On : 17 Jul 2023 17:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 18 04:27:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.613	0.471	23.2	84	0.00
3 P	Chloromethane	0.666	0.605	9.2	103	0.00
4 C	Vinyl Chloride	0.862	0.756	12.3#	97	0.00
5 T	Bromomethane	0.616	0.467	24.2	86	-0.01
6 T	Chloroethane	0.583	0.443	24.0	81	0.00
7 T	Trichlorofluoromethane	1.030	0.846	17.9	90	0.00
8 T	Diethyl Ether	0.343	0.294	14.3	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.467	14.3	91	0.00
10 T	Methyl Iodide	0.749	0.642	14.3	89	0.00
11 T	Tert butyl alcohol	0.103	0.090	12.6	92	0.00
12 CM	1,1-Dichloroethene	0.528	0.436	17.4#	88	0.00
13 T	Acrolein	0.102	0.087	14.7	88	0.00
14 T	Allyl chloride	0.721	0.587	18.6	87	0.00
15 T	Acrylonitrile	0.261	0.235	10.0	93	0.00
16 T	Acetone	0.209	0.179	14.4	94	0.00
17 T	Carbon Disulfide	1.526	1.092	28.4#	79	0.00
18 T	Methyl Acetate	0.884	0.770	12.9	93	0.00
19 T	Methyl tert-butyl Ether	1.738	1.521	12.5	90	0.00
20 T	Methylene Chloride	0.674	0.538	20.2	91	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.502	15.6	89	0.00
22 T	Diisopropyl ether	1.614	1.474	8.7	93	0.00
23 T	Vinyl Acetate	1.026	0.916	10.7	88	0.00
24 P	1,1-Dichloroethane	1.088	0.932	14.3	90	0.00
25 T	2-Butanone	0.329	0.296	10.0	94	0.00
26 T	2,2-Dichloropropane	0.890	0.787	11.6	93	0.00
27 T	cis-1,2-Dichloroethene	0.688	0.609	11.5	93	0.00
28 T	Bromochloromethane	0.512	0.394	23.0	80	0.00
29 T	Tetrahydrofuran	0.216	0.193	10.6	91	0.00
30 C	Chloroform	1.132	1.024	9.5#	93	0.00
31 T	Cyclohexane	0.906	0.772	14.8	97	0.00
32 T	1,1,1-Trichloroethane	1.015	0.978	3.6	99	0.00
33 S	1,2-Dichloroethane-d4	0.670	0.583	13.0	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.339	0.341	-0.6	91	0.00
36 T	1,1-Dichloropropene	0.490	0.530	-8.2	98	0.00
37 T	Ethyl Acetate	0.409	0.426	-4.2	92	0.00
38 T	Carbon Tetrachloride	0.533	0.585	-9.8	98	0.00
39 T	Methylcyclohexane	0.447	0.458	-2.5	91	0.00
40 TM	Benzene	1.452	1.570	-8.1	96	0.00
41 T	Methacrylonitrile	0.208	0.214	-2.9	92	0.00
42 TM	1,2-Dichloroethane	0.494	0.539	-9.1	100	0.00
43 T	Isopropyl Acetate	0.673	0.756	-12.3	99	0.00
44 TM	Trichloroethene	0.391	0.382	2.3	91	0.00
45 C	1,2-Dichloropropane	0.368	0.384	-4.3#	95	0.00
46 T	Dibromomethane	0.259	0.264	-1.9	90	0.00
47 T	Bromodichloromethane	0.519	0.545	-5.0	93	0.00
48 T	Methyl methacrylate	0.312	0.332	-6.4	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
 Data File : VN078548.D
 Acq On : 17 Jul 2023 17:28
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 18 04:27:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	93	0.00
50 S	Toluene-d8	1.282	1.170	8.7	83	0.00
51 T	4-Methyl-2-Pentanone	0.400	0.435	-8.7	94	0.00
52 CM	Toluene	0.900	0.933	-3.7#	91	0.00
53 T	t-1,3-Dichloropropene	0.532	0.560	-5.3	91	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.606	-6.9	91	0.00
55 T	1,1,2-Trichloroethane	0.355	0.379	-6.8	95	0.00
56 T	Ethyl methacrylate	0.496	0.545	-9.9	91	0.00
57 T	1,3-Dichloropropane	0.603	0.620	-2.8	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.177	0.194	-9.6	92	0.00
59 T	2-Hexanone	0.284	0.308	-8.5	91	0.00
60 T	Dibromochloromethane	0.391	0.415	-6.1	91	0.00
61 T	1,2-Dibromoethane	0.365	0.379	-3.8	91	0.00
62 S	4-Bromofluorobenzene	0.395	0.378	4.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.376	0.372	1.1	91	0.00
65 PM	Chlorobenzene	1.118	1.105	1.2	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.430	0.434	-0.9	92	0.00
67 C	Ethyl Benzene	1.889	1.949	-3.2#	91	0.00
68 T	m/p-Xylenes	0.698	0.738	-5.7	91	0.00
69 T	o-Xylene	0.705	0.719	-2.0	91	0.00
70 T	Styrene	1.104	1.194	-8.2	93	0.00
71 P	Bromoform	0.291	0.306	-5.2	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	4.698	4.324	8.0	92	0.00
74 T	N-amyl acetate	1.619	1.461	9.8	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.586	1.405	11.4	95	0.00
76 T	1,2,3-Trichloropropane	1.209	1.134	6.2	94	0.00
77 T	Bromobenzene	1.154	1.010	12.5	90	0.00
78 T	n-propylbenzene	5.024	4.863	3.2	93	0.00
79 T	2-Chlorotoluene	3.291	3.060	7.0	94	0.00
80 T	1,3,5-Trimethylbenzene	3.712	3.552	4.3	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.460	0.474	-3.0	102	0.00
82 T	4-Chlorotoluene	3.070	2.885	6.0	93	0.00
83 T	tert-Butylbenzene	3.152	2.935	6.9	93	0.00
84 T	1,2,4-Trimethylbenzene	3.606	3.556	1.4	96	0.00
85 T	sec-Butylbenzene	3.984	3.801	4.6	94	0.00
86 T	p-Isopropyltoluene	3.199	3.118	2.5	94	0.00
87 T	1,3-Dichlorobenzene	1.804	1.758	2.5	93	0.00
88 T	1,4-Dichlorobenzene	1.817	1.676	7.8	90	0.00
89 T	n-Butylbenzene	2.289	2.400	-4.8	96	0.00
90 T	Hexachloroethane	0.676	0.587	13.2	94	0.00
91 T	1,2-Dichlorobenzene	1.795	1.700	5.3	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.262	0.219	16.4	89	0.00
93 T	1,2,4-Trichlorobenzene	0.573	0.552	3.7	84	0.00
94 T	Hexachlorobutadiene	0.375	0.303	19.2	87	0.00
95 T	Naphthalene	1.958	1.648	15.8	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071723\
Data File : VN078548.D
Acq On : 17 Jul 2023 17:28
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 18 04:27:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.659	0.597	9.4	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6