

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
 Data File : VN082959.D
 Acq On : 17 Jul 2024 13:34
 Operator : JC\MD
 Sample : P3246-17RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CTW2-BL-04RE

Quant Time: Jul 18 05:23:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	91551	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	189300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88084	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	72333	56.362	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.720%
35) Dibromofluoromethane	8.171	113	59761	49.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	10.571	98	256709	57.219	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	114.440%#
62) 4-Bromofluorobenzene	12.853	95	82530	51.211	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	102.420%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	22145	49.011	ug/l	90
20) Methylene Chloride	5.277	84	5550	5.086	ug/l	84
25) 2-Butanone	7.488	43	4223	6.190	ug/l	94
43) Isopropyl Acetate	8.694	43	85016	24.958	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
Data File : VN082959.D
Acq On : 17 Jul 2024 13:34
Operator : JC\MD
Sample : P3246-17RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CTW2-BL-04RE

Quant Time: Jul 18 05:23:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

