

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
 Data File : VN082960.D
 Acq On : 17 Jul 2024 13:58
 Operator : JC\MD
 Sample : P3246-27
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-CTW2-BL-06

Quant Time: Jul 18 05:23:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	108509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	188654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	235136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	87899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	68197	44.835	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.660%
35) Dibromofluoromethane	8.171	113	66477	54.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.820%
50) Toluene-d8	10.571	98	246472	55.125	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.260%
62) 4-Bromofluorobenzene	12.853	95	77614	48.326	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.660%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	16285	30.409	ug/l	95
20) Methylene Chloride	5.289	84	6180	4.778	ug/l #	79
43) Isopropyl Acetate	8.688	43	91166	27.094	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
Data File : VN082960.D
Acq On : 17 Jul 2024 13:58
Operator : JC\MD
Sample : P3246-27
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-CTW2-BL-06

Quant Time: Jul 18 05:23:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

