

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
 Data File : VN082961.D
 Acq On : 17 Jul 2024 14:22
 Operator : JC\MD
 Sample : PB162094TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162094TB

Quant Time: Jul 18 05:24:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	82630	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	172332	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	195310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	83401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	65758	56.771	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.540%
35) Dibromofluoromethane	8.171	113	53232	48.138	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.280%
50) Toluene-d8	10.571	98	208697	51.098	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.853	95	78162	53.276	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	106.560%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	10211	25.039	ug/l	91
20) Methylene Chloride	5.277	84	3725	3.782	ug/l	87
43) Isopropyl Acetate	8.694	43	110426	36.955	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
Data File : VN082961.D
Acq On : 17 Jul 2024 14:22
Operator : JC\MD
Sample : PB162094TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162094TB

Quant Time: Jul 18 05:24:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

