

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
 Data File : VN082962.D
 Acq On : 17 Jul 2024 14:47
 Operator : JC\MD
 Sample : PB162095TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB162095TB

Quant Time: Jul 18 05:24:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	102509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	212345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	183899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	70617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	83245	57.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.860%
35) Dibromofluoromethane	8.165	113	68646	50.380	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.760%
50) Toluene-d8	10.570	98	256219	50.912	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.820%
62) 4-Bromofluorobenzene	12.853	95	63258	34.993	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	69.980%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	13861	27.398	ug/l	96
20) Methylene Chloride	5.283	84	8328	6.816	ug/l	86
43) Isopropyl Acetate	8.694	43	166122	45.815	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
Data File : VN082962.D
Acq On : 17 Jul 2024 14:47
Operator : JC\MD
Sample : PB162095TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB162095TB

Quant Time: Jul 18 05:24:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

