

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
 Data File : VN082967.D
 Acq On : 17 Jul 2024 16:48
 Operator : JC\MD
 Sample : P3266-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-1

Quant Time: Jul 18 05:27:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	73271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	155383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	167311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	65022	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	58831	57.278	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.560%
35) Dibromofluoromethane	8.171	113	48404	48.547	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.100%
50) Toluene-d8	10.571	98	185816	50.458	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.853	95	58277	44.055	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	88.120%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	9818	27.150	ug/l	92
20) Methylene Chloride	5.265	84	5109	5.850	ug/l #	70
43) Isopropyl Acetate	8.694	43	109801	41.078	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071724\
Data File : VN082967.D
Acq On : 17 Jul 2024 16:48
Operator : JC\MD
Sample : P3266-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-1

Quant Time: Jul 18 05:27:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

