

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071725\
 Data File : VN087339.D
 Acq On : 17 Jul 2025 10:56
 Operator : JC\MD
 Sample : VN0717WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0717WBL01

Quant Time: Jul 18 01:21:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	171751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	343640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	297458	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	124823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	160879	55.204	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.400%	
35) Dibromofluoromethane	8.153	113	119450	50.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.780%	
50) Toluene-d8	10.547	98	429548	50.800	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	12.829	95	134801	43.151	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 86.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071725\
Data File : VN087339.D
Acq On : 17 Jul 2025 10:56
Operator : JC\MD
Sample : VN0717WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0717WBL01

Quant Time: Jul 18 01:21:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

