

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049873.D
 Acq On : 18 Jul 2018 15:27
 Operator : MD\SY
 Sample : PB111206ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111206ZHE#07

Quant Time: Jul 19 04:16:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	582588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	927622	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	812007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	301846	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	406969	50.82	ug/l	0.00
Spiked Amount			Recovery	=	101.64%	
35) Dibromofluoromethane	7.59	113	359027	46.10	ug/l	0.00
Spiked Amount			Recovery	=	92.20%	
50) Toluene-d8	10.09	98	1277701	45.03	ug/l	0.00
Spiked Amount			Recovery	=	90.06%	
62) 4-Bromofluorobenzene	12.40	95	358182	36.65	ug/l	0.00
Spiked Amount			Recovery	=	73.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	23508	11.34	ug/l	97
18) Methyl Acetate	4.33	43	13227	1.23	ug/l	95
20) Methylene Chloride	4.55	84	13071	1.66	ug/l #	92
43) Isopropyl Acetate	8.17	43	250465	20.24	ug/l	100
59) 2-Hexanone	10.77	43	23930	5.15	ug/l #	60
95) Naphthalene	15.14	128	7498	4.97	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049873.D
Acq On : 18 Jul 2018 15:27
Operator : MD\SY
Sample : PB111206ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111206ZHE#07

Quant Time: Jul 19 04:16:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

