

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049875.D
 Acq On : 18 Jul 2018 16:19
 Operator : MD\SY
 Sample : PB111206ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111206ZHE#09

Quant Time: Jul 19 04:27:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	561963	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	925044	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	767988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	273280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	399429	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
35) Dibromofluoromethane	7.59	113	352350	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
50) Toluene-d8	10.09	98	1230722	43.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.00%	
62) 4-Bromofluorobenzene	12.40	95	329265	33.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.58%	
Target Compounds						
16) Acetone	3.83	43	23338	11.67	ug/l	95
18) Methyl Acetate	4.34	43	14969	1.58	ug/l #	83
20) Methylene Chloride	4.55	84	14449	1.90	ug/l	95
43) Isopropyl Acetate	8.17	43	238927	19.36	ug/l	98
59) 2-Hexanone	10.77	43	21652	4.67	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049875.D
Acq On : 18 Jul 2018 16:19
Operator : MD\SY
Sample : PB111206ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111206ZHE#09

Quant Time: Jul 19 04:27:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

