

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049883.D
 Acq On : 18 Jul 2018 19:47
 Operator : MD\SY
 Sample : PB111206ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111206ZHE#15

Quant Time: Jul 19 05:23:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	553272	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	946442	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	792643	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	296991	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	412510	54.25	ug/l	0.00
Spiked Amount			Recovery	=	108.50%	
35) Dibromofluoromethane	7.59	113	356845	44.91	ug/l	0.00
Spiked Amount			Recovery	=	89.82%	
50) Toluene-d8	10.09	98	1250149	43.18	ug/l	0.00
Spiked Amount			Recovery	=	86.36%	
62) 4-Bromofluorobenzene	12.40	95	349099	35.01	ug/l	0.00
Spiked Amount			Recovery	=	70.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	22998	11.68	ug/l	95
18) Methyl Acetate	4.34	43	23110	2.96	ug/l	99
20) Methylene Chloride	4.55	84	13976	1.87	ug/l	94
43) Isopropyl Acetate	8.17	43	309366	24.50	ug/l	93
59) 2-Hexanone	10.77	43	21252	4.48	ug/l #	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049883.D
Acq On : 18 Jul 2018 19:47
Operator : MD\SY
Sample : PB111206ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111206ZHE#15

Quant Time: Jul 19 05:23:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

