

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049885.D
 Acq On : 18 Jul 2018 20:39
 Operator : MD\SY
 Sample : PB111206ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111206ZHE#17

Quant Time: Jul 19 05:46:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	554390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	937061	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	765968	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	260213	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	400947	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	7.59	113	351654	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	
50) Toluene-d8	10.09	98	1233119	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	
62) 4-Bromofluorobenzene	12.40	95	318248	32.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.48%	
Target Compounds						
16) Acetone	3.83	43	18972	9.62	ug/l	97
18) Methyl Acetate	4.33	43	26919	3.58	ug/l	98
20) Methylene Chloride	4.55	84	10897	1.45	ug/l	93
43) Isopropyl Acetate	8.17	43	329016	26.32	ug/l #	94
59) 2-Hexanone	10.77	43	22431	4.78	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049885.D
Acq On : 18 Jul 2018 20:39
Operator : MD\SY
Sample : PB111206ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111206ZHE#17

Quant Time: Jul 19 05:46:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

