

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049892.D
 Acq On : 19 Jul 2018 00:33
 Operator : MD\SY
 Sample : PB111206ZHE#19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111206ZHE#19

Quant Time: Jul 19 06:51:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	565663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	947477	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	788173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	275323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	408431	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
35) Dibromofluoromethane	7.59	113	360667	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
50) Toluene-d8	10.09	98	1272466	43.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.82%	
62) 4-Bromofluorobenzene	12.40	95	336454	33.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.42%	
Target Compounds						
16) Acetone	3.83	43	20290	10.08	ug/l	93
18) Methyl Acetate	4.34	43	22654	2.80	ug/l	95
20) Methylene Chloride	4.55	84	9658	1.26	ug/l	97
43) Isopropyl Acetate	8.17	43	270133	21.37	ug/l	98
59) 2-Hexanone	10.77	43	23155	4.88	ug/l #	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049892.D
Acq On : 19 Jul 2018 00:33
Operator : MD\SY
Sample : PB111206ZHE#19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111206ZHE#19

Quant Time: Jul 19 06:51:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

