

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049896.D
 Acq On : 19 Jul 2018 2:17
 Operator : MD\SY
 Sample : J3357-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELDBLANK_71718

Manual Integrations
 APPROVED

apatel
 7/19/2018 4:30:07 PM

Quant Time: Jul 19 14:12:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	553149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	926019	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	754316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	277771	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	401736	52.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.68%	
35) Dibromofluoromethane	7.59	113	359671	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	10.09	98	1228297	43.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.72%	
62) 4-Bromofluorobenzene	12.40	95	330713	33.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049896.D
Acq On : 19 Jul 2018 2:17
Operator : MD\SY
Sample : J3357-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELDBLANK_71718

Manual Integrations
APPROVED

apatel
7/19/2018 4:30:07 PM

Quant Time: Jul 19 14:12:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

