

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049897.D
 Acq On : 19 Jul 2018 2:43
 Operator : MD\SY
 Sample : J3357-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW3

Quant Time: Jul 19 14:13:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 13:09:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	552638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	912820	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	747186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	280805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	400140	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
35) Dibromofluoromethane	7.59	113	358036	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
50) Toluene-d8	10.09	98	1212239	43.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.84%	
62) 4-Bromofluorobenzene	12.40	95	325723	33.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.74%	
Target Compounds						
3) Chloromethane	2.06	50	9987	1.32	ug/l	Qvalue 100
30) Chloroform	7.37	83	765821	60.29	ug/l	99
47) Bromodichloromethane	9.41	83	6756	0.63	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049897.D
Acq On : 19 Jul 2018 2:43
Operator : MD\SY
Sample : J3357-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW3

Quant Time: Jul 19 14:13:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 13:09:48 2018
Response via : Initial Calibration

