


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
Data File : VN049862.D
Acq On : 18 Jul 2018 8:58
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Title : SW846 8260
Last Update : Thu Jul 19 16:34:06 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3349

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	6664	PASS
75	95	30	60	49.8	18024	PASS
95	95	100	100	100.0	36160	PASS
96	95	5	9	6.9	2493	PASS
173	174	0.00	2	1.8	612	PASS
174	95	50	100	93.0	33632	PASS
175	174	5	9	7.4	2476	PASS
176	174	95	101	99.8	33573	PASS
177	176	5	9	7.1	2368	PASS