


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071818\
 Data File : VN049886.D
 Acq On : 18 Jul 2018 21:05
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Title : SW846 8260
 Last Update : Thu Jul 19 01:58:39 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3348

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	10974	PASS
75	95	30	60	49.3	27573	PASS
95	95	100	100	100.0	55968	PASS
96	95	5	9	6.5	3659	PASS
173	174	0.00	2	0.9	483	PASS
174	95	50	100	92.1	51554	PASS
175	174	5	9	7.9	4064	PASS
176	174	95	101	95.7	49357	PASS
177	176	5	9	6.5	3215	PASS