

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071819\
 Data File : VN056807.D
 Acq On : 18 Jul 2019 12:05
 Operator : JC/SP
 Sample : K3882-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 U3MW-5S

Quant Time: Jul 19 09:08:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071319W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jul 12 23:58:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	48938	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	266541	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	212718	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	99839	30.12	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.40%
60) 4-Bromofluorobenzene	12.40	95	74486	22.88	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.27%
63) Toluene-d8	10.09	98	304172	31.30	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.33%

Target Compounds

					Ovalue	
15) Acetone	3.83	58	1695	2.885	ug/l	# 51

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071819\
Data File : VN056807.D
Acq On : 18 Jul 2019 12:05
Operator : JC/SP
Sample : K3882-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
U3MW-5S

Quant Time: Jul 19 09:08:29 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N071319W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jul 12 23:58:13 2019
Response via : Initial Calibration

