

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071819\
 Data File : VN056814.D
 Acq On : 18 Jul 2019 16:04
 Operator : JC/SP
 Sample : K3892-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-1

Quant Time: Jul 19 09:19:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 07:03:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	254812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	460913	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	477124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	157310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	150385	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	7.59	113	138008	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	10.09	98	580546	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	12.40	95	202310	54.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.30%	
Target Compounds						
3) Chloromethane	2.06	50	4375	1.360	ug/l	91
16) Acetone	3.82	43	61684	47.491	ug/l	99
30) Chloroform	7.38	83	16036	3.380	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071819\
Data File : VN056814.D
Acq On : 18 Jul 2019 16:04
Operator : JC/SP
Sample : K3892-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-1

Quant Time: Jul 19 09:19:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071519W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 07:03:34 2019
Response via : Initial Calibration

