

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
 Data File : VN073495.D
 Acq On : 18 Jul 2022 13:37
 Operator : JC\MD
 Sample : N3711-01
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PSC124019

Quant Time: Jul 18 15:31:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	251378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	430415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	426581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	214394	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	164063	45.210	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.420%
35) Dibromofluoromethane	7.945	113	135748	47.404	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	10.374	98	462900	45.514	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.674	95	212644	54.405	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
Data File : VN073495.D
Acq On : 18 Jul 2022 13:37
Operator : JC\MD
Sample : N3711-01
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PSC124019

Quant Time: Jul 18 15:31:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

