

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
 Data File : VN073496.D
 Acq On : 18 Jul 2022 14:02
 Operator : JC\MD
 Sample : N3711-03
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PSC124042

Quant Time: Jul 18 15:31:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	256748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	431090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	420526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	208692	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	166830	45.011	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.020%
35) Dibromofluoromethane	7.945	113	135273	47.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	10.380	98	470719	46.210	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.420%
62) 4-Bromofluorobenzene	12.674	95	212815	54.364	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
Data File : VN073496.D
Acq On : 18 Jul 2022 14:02
Operator : JC\MD
Sample : N3711-03
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PSC124042

Quant Time: Jul 18 15:31:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

