

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
 Data File : VN073503.D
 Acq On : 18 Jul 2022 16:53
 Operator : JC\MD
 Sample : N3731-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Jul 19 07:05:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	238227	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	424106	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	400395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	189858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	162614	47.284	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.560%
35) Dibromofluoromethane	7.951	113	135950	48.181	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.360%
50) Toluene-d8	10.381	98	454719	45.374	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.740%
62) 4-Bromofluorobenzene	12.675	95	193043	50.125	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.260%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	43675	22.625	ug/l	100
18) Methyl Acetate	4.799	43	5915	1.093	ug/l #	80
20) Methylene Chloride	5.010	84	24429	7.201	ug/l	92
37) Ethyl Acetate	7.340	43	50444	7.689	ug/l	98
43) Isopropyl Acetate	8.498	43	199818	21.194	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
Data File : VN073503.D
Acq On : 18 Jul 2022 16:53
Operator : JC\MD
Sample : N3731-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Jul 19 07:05:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

