

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
 Data File : VN073504.D
 Acq On : 18 Jul 2022 17:18
 Operator : JC\MD
 Sample : N3731-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Jul 19 07:05:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	245543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	436519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	410382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	190426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	171104	48.271	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.540%
35) Dibromofluoromethane	7.951	113	142645	49.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.240%
50) Toluene-d8	10.381	98	462506	44.839	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.680%
62) 4-Bromofluorobenzene	12.675	95	191812	48.389	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.780%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	93061	46.772	ug/l	94
18) Methyl Acetate	4.793	43	6614	1.186	ug/l #	66
20) Methylene Chloride	5.016	84	23127	6.614	ug/l #	81
43) Isopropyl Acetate	8.498	43	21597	2.226	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
Data File : VN073504.D
Acq On : 18 Jul 2022 17:18
Operator : JC\MD
Sample : N3731-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Jul 19 07:05:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

