

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
 Data File : VN073506.D
 Acq On : 18 Jul 2022 18:07
 Operator : JC\MD
 Sample : N3760-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1821

Quant Time: Jul 19 07:05:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/19/2022
 Supervised By : Mahesh Dadoda 07/19/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	249008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	426308	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	391285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	169929	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	169293	47.095	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.200%		
35) Dibromofluoromethane	7.957	113	140479	49.529	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.060%		
50) Toluene-d8	10.380	98	453252	44.994	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.980%		
62) 4-Bromofluorobenzene	12.674	95	179111	46.267	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.540%		
Target Compounds					Qvalue	
16) Acetone	4.234	43	8826	4.374	ug/l	94
20) Methylene Chloride	5.010	84	2370	0.668	ug/l #	88
25) 2-Butanone	7.263	43	4837m	1.506	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071822\
Data File : VN073506.D
Acq On : 18 Jul 2022 18:07
Operator : JC\MD
Sample : N3760-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1821

Quant Time: Jul 19 07:05:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 07/19/2022
Supervised By :Mahesh Dadoda 07/19/2022

