

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071823\
 Data File : VN078551.D
 Acq On : 18 Jul 2023 11:03
 Operator : JC\MD
 Sample : VN0718MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0718MBL01

Quant Time: Jul 19 01:44:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	406941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	714817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	641541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	220434	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	258263	47.340	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.680%
35) Dibromofluoromethane	8.177	113	270537	55.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.500%
50) Toluene-d8	10.570	98	818157	44.633	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.260%
62) 4-Bromofluorobenzene	12.853	95	276153	48.909	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071823\
Data File : VN078551.D
Acq On : 18 Jul 2023 11:03
Operator : JC\MD
Sample : VN0718MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0718MBL01

Quant Time: Jul 19 01:44:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

