

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071823\
 Data File : VN078555.D
 Acq On : 18 Jul 2023 12:49
 Operator : JC\MD
 Sample : PB154251TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154251TB

Quant Time: Jul 19 01:47:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	399520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	724270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	694759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	226392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	266771	49.808	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.620%
35) Dibromofluoromethane	8.177	113	266251	54.154	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.300%
50) Toluene-d8	10.570	98	883805	47.585	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.160%
62) 4-Bromofluorobenzene	12.853	95	294120	51.411	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.820%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	116725	69.963	ug/l	98
20) Methylene Chloride	5.294	84	23870	3.860	ug/l	92
25) 2-Butanone	7.500	43	84089	32.018	ug/l #	84
37) Ethyl Acetate	7.577	43	24636	4.163	ug/l	98
43) Isopropyl Acetate	8.694	43	267132	27.417	ug/l #	85
51) 4-Methyl-2-Pentanone	10.453	43	292656	50.499	ug/l	91
58) 2-Chloroethyl Vinyl ether	10.171	63	5230	2.044	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071823\
Data File : VN078555.D
Acq On : 18 Jul 2023 12:49
Operator : JC\MD
Sample : PB154251TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154251TB

Quant Time: Jul 19 01:47:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

