

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071824\
 Data File : VN082973.D
 Acq On : 18 Jul 2024 13:27
 Operator : JC\MD
 Sample : VN0718WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0718WBL01

Quant Time: Jul 19 00:42:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	73281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	145862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	144948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	57745	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	59223	57.652	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.300%
35) Dibromofluoromethane	8.165	113	49259	52.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.260%
50) Toluene-d8	10.570	98	171692	49.666	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.847	95	53025	42.701	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	85.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071824\
Data File : VN082973.D
Acq On : 18 Jul 2024 13:27
Operator : JC\MD
Sample : VN0718WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0718WBL01

Quant Time: Jul 19 00:42:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

