

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049941.D
 Acq On : 20 Jul 2018 2:48
 Operator : MD\SY
 Sample : J4035-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-2

Quant Time: Jul 20 05:21:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	555458	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	963162	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	810074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	326325	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	402895	52.77	ug/l	0.00
Spiked Amount			Recovery	=	105.54%	
35) Dibromofluoromethane	7.59	113	360000	44.52	ug/l	0.00
Spiked Amount			Recovery	=	89.04%	
50) Toluene-d8	10.09	98	1282334	43.53	ug/l	0.00
Spiked Amount			Recovery	=	87.06%	
62) 4-Bromofluorobenzene	12.40	95	365852	36.06	ug/l	0.00
Spiked Amount			Recovery	=	72.12%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.85	85	30	Below Cal	#	44
5) Bromomethane	2.57	94	547	Below Cal	#	30
14) Allyl chloride	4.55	41	88195	9.29	ug/l	25
16) Acetone	3.83	43	22221	11.24	ug/l	96
18) Methyl Acetate	4.34	43	26879	3.56	ug/l	96
20) Methylene Chloride	4.55	84	2560832	340.67	ug/l	99
25) 2-Butanone	6.85	43	3814	1.35	ug/l	82
29) Tetrahydrofuran	7.22	42	1901	1.07	ug/l	64
31) Cyclohexane	7.67	56	10343	1.02	ug/l	1
36) 1,1-Dichloropropene	7.67	75	40082	3.74	ug/l	48
38) Carbon Tetrachloride	7.67	117	51861	4.52	ug/l	16
43) Isopropyl Acetate	8.17	43	396889	30.89	ug/l	86
48) Methyl methacrylate	9.18	41	8017	1.24	ug/l	21
51) 4-Methyl-2-Pentanone	9.99	43	2574	0.36	ug/l	91
56) Ethyl methacrylate	10.26	69	265	2.18	ug/l	26
59) 2-Hexanone	10.77	43	19070	3.95	ug/l	56
68) m/p-Xylenes	11.62	106	4737	0.38	ug/l	96
71) Bromoform	12.40	173	2822	0.50	ug/l	1
76) 1,2,3-Trichloropropane	12.40	75	178872	29.22	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.40	75	178872	91.37	ug/l	13
84) 1,2,4-Trimethylbenzene	13.04	105	7685	0.38	ug/l	92
85) sec-Butylbenzene	13.04	105	7685	0.33	ug/l	56
90) Hexachloroethane	14.00	117	1279	0.29	ug/l	4
92) 1,2-Dibromo-3-Chloropropan	14.26	75	341	0.30	ug/l	1
93) 1,2,4-Trichlorobenzene	14.90	180	394	3.53	ug/l	74
95) Naphthalene	15.14	128	1800349	110.51	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049941.D
Acq On : 20 Jul 2018 2:48
Operator : MD\SY
Sample : J4035-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-2

Quant Time: Jul 20 05:21:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration






#1




Pentafluorobenzene
 Concen: 50.00 ug/l
 RT: 7.67 min Scan# 1891
 Delta R.T. 0.00 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48

Instrument :
 MSVOA_N
 ClientSampled :
 MH-2

Tot Ion:168 Resp: 555458
 Ion Ratio Lower Upper
 168 100
 99 50.2 40.6 61.0




Abundance Ion 167.90 (167.60 to 168.60): V
 Ion 98.90 (98.60 to 99.60): VNO





#2


Dichlorodifluoromethane
 Concen: Below Cal
 RT: 1.85 min Scan# 82
 Delta R.T. 0.00 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48

Tgt Ion: 85 Resp: 30
 Ion Ratio Lower Upper
 85 100
 87 0.0 15.4 46.1#



Abundance Ion 84.90 (84.60 to 85.60): VNO
 Ion 86.90 (86.60 to 87.60): VNO





#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 306

Delta R.T. 0.01 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :


MH-2

Tot Ion: 94 Resp: 547

Ion Ratio Lower Upper




94 100

96 26.6 75.7 113.5#



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0



#14

Allyl chloride

Concen: 9.29 ug/l

RT: 4.55 min Scan# 921

Delta R.T. 0.23 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48


Tgt Ion: 41 Resp: 88195

Ion Ratio Lower Upper

41 100

39 0.9 53.2 79.8#



76 0.0 29.0 43.6#




Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0

Ion 75.90 (75.60 to 76.60): VN0





#16

Acetone

Concen: 11.24 ug/l

RT: 3.83 min Scan# 696

Delta R.T. 0.01 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :





MH-2

Tot Ion: 43 Resp: 22221

Ion Ratio Lower Upper

43 100

58 34.4 25.9 38.9



#18

Methyl Acetate

Concen: 3.56 ug/l

RT: 4.34 min Scan# 854

Delta R.T. 0.01 min

Lab File: VN049941.D




Acq: 20 Jul 2018 2:48


Tgt Ion: 43 Resp: 26879

Ion Ratio Lower Upper

43 100

74 22.8 19.8 29.8









#20
Methvlene Chloride
Concen: 340.67 ug/l
RT: 4.55 min Scan# 921
Delta R.T. 0.00 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48

Instrument :
MSVOA_N
ClientSampled :
MH-2

Tot Ion: 84 Resp: 2560832




Ion	Ratio	Lower	Upper
84	100		
49	120.2	95.4	143.2
51	36.8	28.4	42.6
86	65.3	52.6	79.0




#25
2-Butanone
Concen: 1.35 ug/l
RT: 6.85 min Scan# 1637
Delta R.T. 0.01 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48

Tgt Ion: 43 Resp: 3814

Ion	Ratio	Lower	Upper
43	100		
72	18.4	22.1	33.1#





#29

Tetrahydrofuran

Concen: 1.07 ug/l

RT: 7.22 min Scan# 1752

Delta R.T. 0.01 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :

MH-2





Tot Ion: 42 Resp: 1901

Ion Ratio Lower Upper

42 100

72 25.8 35.6 53.4#

71 15.4 34.0 51.0#



#31

Cyclohexane

Concen: 1.02 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.02 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48




Tgt Ion: 56 Resp: 10343


Ion Ratio Lower Upper

56 100

69 180.2 26.6 40.0#

84 135.8 69.0 103.6#









#33
 1,2-Dichloroethane-d4
 Concen: 52.77 ug/l
 RT: 8.03 min Scan# 2003
 Delta R.T. 0.00 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48




Instrument :
 MSVOA_N
 ClientSampled :
 MH-2


Tot Ion: 65 Resp: 402895
 Ion Ratio Lower Upper
 65 100
 67 53.2 0.0 107.2



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. -0.00 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48

Tgt Ion: 114 Resp: 963162
 Ion Ratio Lower Upper
 114 100
 63 20.2 0.0 39.8
 88 15.2 0.0 30.6





#35

Dibromofluoromethane

Concen: 44.52 ug/l

RT: 7.59 min Scan# 1867

Delta R.T. -0.00 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampleId :

MH-2


Tot Ion:113 Resp: 360000

Ion Ratio Lower Upper

113 100

111 101.5 81.8 122.8




192 21.5 17.8 26.6



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 3.74 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.13 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48


Tgt Ion: 75 Resp: 40082

Ion Ratio Lower Upper

75 100

110 7.8 18.1 54.4#



77 0.0 24.9 37.3#




Abundance Ion 74.90 (74.60 to 75.60): V

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): V









#38
Carbon Tetrachloride
Concen: 4.52 ug/l
RT: 7.67 min Scan# 1891
Delta R.T. -0.10 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48




Instrument :
MSVOA_N
ClientSampleId :
MH-2


Tot Ion:117 Resp: 51861
Ion Ratio Lower Upper
117 100
119 5.8 77.2 115.8#
121 0.0 25.0 37.6#



#43
Isopropyl Acetate
Concen: 30.89 ug/l
RT: 8.17 min Scan# 2047
Delta R.T. 0.00 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48

Tgt Ion: 43 Resp: 396889
Ion Ratio Lower Upper
43 100
61 11.9 14.9 22.3#
87 7.5 10.0 15.0#





#48

Methvl methacrylate

Concen: 1.24 ug/l

RT: 9.18 min Scan# 2362

Delta R.T. -0.02 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampleId :

MH-2


Tot Ion: 41 Resp: 8017

Ion Ratio Lower Upper

41 100

69 0.0 71.5 107.3#

39 16.7 44.7 67.1#




Abundance Ion 41.00 (40.70 to 41.70): VN049866.D (-2355) (-)


Ion 69.00 (68.70 to 69.70): VN049866.D (-2355) (-)

Ion 39.00 (38.70 to 39.70): VN049866.D (-2355) (-)

Time-->



Time-->



#50

Toluene-d8

Concen: 43.53 ug/l

RT: 10.09 min Scan# 2645

Delta R.T. 0.00 min

Lab File: VN049941.D


Acq: 20 Jul 2018 2:48

Tgt Ion: 98 Resp: 1282334

Ion Ratio Lower Upper

98 100


100 63.2 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN049866.D (-2626) (-)


Ion 100.00 (99.70 to 100.70): VN049866.D (-2626) (-)

Time-->



Time-->

Time-->



#51

4-Methyl-2-Pentanone

Concen: 0.36 ug/l

RT: 9.99 min Scan# 2612

Delta R.T. 0.00 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :





MH-2

Tot Ion: 43 Resp: 2574

Ion Ratio Lower Upper

43 100

58 34.0 31.8 47.8



#56

Ethyl methacrylate

Concen: 2.18 ug/l

RT: 10.26 min Scan# 2697

Delta R.T. -0.17 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48




Tgt Ion: 69 Resp: 265


Ion Ratio Lower Upper

69 100

41 0.0 51.6 77.4#

39 0.0 25.8 38.8#





#59

2-Hexanone

Concen: 3.95 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :





MH-2

Tot Ion: 43 Resp: 19070

Ion Ratio Lower Upper

43 100

58 23.5 27.8 83.4#



#62

4-Bromofluorobenzene

Concen: 36.06 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.00 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48




Tgt Ion: 95 Resp: 365852


Ion Ratio Lower Upper

95 100

174 90.7 0.0 179.6

176 87.3 0.0 174.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. -0.00 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48




Instrument :
MSVOA_N
ClientSampled :
MH-2


Tot Ion:117 Resp: 810074
Ion Ratio Lower Upper
117 100
82 54.0 41.9 62.9
119 32.4 25.4 38.2



#68
m/p-Xylenes
Concen: 0.38 ug/l
RT: 11.62 min Scan# 3119
Delta R.T. -0.00 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48

Tgt Ion:106 Resp: 4737
Ion Ratio Lower Upper
106 100
91 198.8 163.4 245.2







#71
Bromoform
Concen: 0.50 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.28 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48

Instrument :
MSVOA_N
ClientSampleId :
MH-2



Tot Ion:173 Resp: 2822
Ion Ratio Lower Upper
173 100
175 703.5 24.6 73.8#
254 0.0 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V




Time--> 12.36 12.38 12.40 12.42




#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.35 min Scan# 3656
Delta R.T. -0.00 min
Lab File: VN049941.D
Acq: 20 Jul 2018 2:48


Tgt Ion:152 Resp: 326325
Ion Ratio Lower Upper
152 100
115 56.0 28.1 84.2
150 158.2 0.0 347.6




Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V



Time--> 13.25 13.30 13.35 13.40





#76

1,2,3-Trichloropropane

Concen: 29.22 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :





MH-2

Tot Ion: 75 Resp: 178872

Ion Ratio Lower Upper

75 100

77 1.3 0.0 0.0#



#81

trans-1,4-Dichloro-2-butene

Concen: 91.37 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48




Tgt Ion: 75 Resp: 178872


Ion Ratio Lower Upper

75 100

53 0.3 73.9 110.9#

89 0.0 35.9 53.9#





#84

1,2,4-Trimethylbenzene

Concen: 0.38 ug/l

RT: 13.04 min Scan# 3562

Delta R.T. 0.00 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :


MH-2

Tot Ion:105 Resp: 7685

Ion Ratio Lower Upper




105 100

120 40.1 22.7 68.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#85

sec-Butylbenzene

Concen: 0.33 ug/l

RT: 13.04 min Scan# 3562

Delta R.T. -0.13 min

Lab File: VN049941.D


Acq: 20 Jul 2018 2:48

Tgt Ion:105 Resp: 7685

Ion Ratio Lower Upper



105 100


134 0.0 10.1 30.1#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V





#90

Hexachloroethane

Concen: 0.29 ug/l

RT: 14.00 min Scan# 3859

Delta R.T. 0.12 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48

Instrument :

MSVOA_N

ClientSampled :


MH-2

Tot Ion:117 Resp: 1279

Ion Ratio Lower Upper




117 100

201 0.0 45.3 135.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 0.30 ug/l

RT: 14.26 min Scan# 3940

Delta R.T. -0.01 min

Lab File: VN049941.D

Acq: 20 Jul 2018 2:48


Tgt Ion: 75 Resp: 341

Ion Ratio Lower Upper

75 100

155 0.0 45.7 137.1#



157 0.0 56.9 170.7#




Abundance Ion 74.90 (74.60 to 75.60): V

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 3.53 ug/l
 RT: 14.90 min Scan# 4140
 Delta R.T. -0.01 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-2

Tot Ion:180 Resp: 394
 Ion Ratio Lower Upper
 180 100
 182 114.0 47.3 141.9
 145 53.3 14.4 43.0#



#95
 Naphthalene
 Concen: 110.51 ug/l
 RT: 15.14 min Scan# 4213
 Delta R.T. 0.00 min
 Lab File: VN049941.D
 Acq: 20 Jul 2018 2:48

Tgt Ion:128 Resp: 1800349
 Ion Ratio Lower Upper
 128 100
 127 13.1 10.4 15.6
 129 10.8 8.6 13.0

