

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049946.D
 Acq On : 20 Jul 2018 4:58
 Operator : MD\SY
 Sample : J4030-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB6-12-7

Quant Time: Jul 20 05:23:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	566587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	955474	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	791336	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	288754	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	408294	52.43	ug/l	0.00
Spiked Amount			Recovery	=	104.86%	
35) Dibromofluoromethane	7.59	113	358761	44.73	ug/l	0.00
Spiked Amount			Recovery	=	89.46%	
50) Toluene-d8	10.09	98	1273042	43.56	ug/l	0.00
Spiked Amount			Recovery	=	87.12%	
62) 4-Bromofluorobenzene	12.40	95	345639	34.34	ug/l	0.00
Spiked Amount			Recovery	=	68.68%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	2138	0.28	ug/l	93
5) Bromomethane	2.57	94	271	Below Cal	#	38
8) Diethyl Ether	3.41	74	3420	0.86	ug/l	99
11) Tert butyl alcohol	4.80	59	865	1.56	ug/l	# 72
16) Acetone	3.83	43	26026	12.91	ug/l	99
18) Methyl Acetate	4.33	43	28402	3.72	ug/l	98
20) Methylene Chloride	4.55	84	75796	9.89	ug/l	95
25) 2-Butanone	6.85	43	6318	2.19	ug/l	94
28) Bromochloromethane	7.38	49	1169	0.20	ug/l	# 11
29) Tetrahydrofuran	7.22	42	1737	0.96	ug/l	# 75
30) Chloroform	7.37	83	15102	1.16	ug/l	99
31) Cyclohexane	7.67	56	10439	1.01	ug/l	# 1
36) 1,1-Dichloropropene	7.67	75	40284	3.79	ug/l	# 48
38) Carbon Tetrachloride	7.67	117	51782	4.55	ug/l	# 16
43) Isopropyl Acetate	8.17	43	206280	16.18	ug/l	95
47) Bromodichloromethane	9.40	83	4585	0.41	ug/l	# 92
48) Methyl methacrylate	9.18	41	8443	1.32	ug/l	# 22
51) 4-Methyl-2-Pentanone	9.99	43	2443	0.34	ug/l	98
56) Ethyl methacrylate	10.27	69	222	2.18	ug/l	# 1
59) 2-Hexanone	10.77	43	8980	1.88	ug/l	# 57
68) m/p-Xylenes	11.62	106	2826	0.23	ug/l	98
71) Bromoform	12.40	173	3039	0.55	ug/l	# 1
76) 1,2,3-Trichloropropane	12.40	75	168635	31.13	ug/l	# 100
81) trans-1,4-Dichloro-2-buten	12.40	75	168635	97.35	ug/l	# 13
95) Naphthalene	15.14	128	69934	9.15	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049946.D
Acq On : 20 Jul 2018 4:58
Operator : MD\SY
Sample : J4030-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB6-12-7

Quant Time: Jul 20 05:23:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :


SB6-12-7

Tot Ion:168 Resp: 566587

Ion Ratio Lower Upper




168 100

99 51.6 40.6 61.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#3

Chloromethane

Concen: 0.28 ug/l

RT: 2.06 min Scan# 145

Delta R.T. -0.00 min

Lab File: VN049946.D


Acq: 20 Jul 2018 4:58

Tgt Ion: 50 Resp: 2138

Ion Ratio Lower Upper



50 100


52 28.4 26.1 39.1



Abundance Ion 49.90 (49.60 to 50.60): VN0

Ion 51.90 (51.60 to 52.60): VN0





#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 305

Delta R.T. 0.01 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampled :





SB6-12-7

Tot Ion: 94 Resp: 271

Ion Ratio Lower Upper

94 100

96 34.9 75.7 113.5#



#8

Diethyl Ether

Concen: 0.86 ug/l

RT: 3.41 min Scan# 567

Delta R.T. 0.00 min

Lab File: VN049946.D




Acq: 20 Jul 2018 4:58


Tgt Ion: 74 Resp: 3420

Ion Ratio Lower Upper

74 100

45 97.4 48.4 145.2









#11

Tert butyl alcohol
 Concen: 1.56 ug/l
 RT: 4.80 min Scan# 997
 Delta R.T. -0.01 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58

Instrument :
 MSVOA_N
 ClientSampled :
 SB6-12-7




Tot Ion: 59 Resp: 865
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.5 12.7#




#16

Acetone
 Concen: 12.91 ug/l
 RT: 3.83 min Scan# 696
 Delta R.T. 0.01 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58

Tgt Ion: 43 Resp: 26026
 Ion Ratio Lower Upper
 43 100
 58 33.2 25.9 38.9









#18
Methvl Acetate
Concen: 3.72 ug/l
RT: 4.33 min Scan# 853
Delta R.T. 0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58




Instrument :
MSVOA_N
ClientSampled :
SB6-12-7


Tot Ion: 43 Resp: 28402
Ion Ratio Lower Upper
43 100
74 23.7 19.8 29.8



#20
Methylene Chloride
Concen: 9.89 ug/l
RT: 4.55 min Scan# 920
Delta R.T. -0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58

Tgt Ion: 84 Resp: 75796
Ion Ratio Lower Upper
84 100
49 111.0 95.4 143.2
51 35.1 28.4 42.6
86 63.3 52.6 79.0









#25
2-Butanone
Concen: 2.19 ug/l
RT: 6.85 min Scan# 1637
Delta R.T. 0.01 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58




Instrument :
MSVOA_N
ClientSampleId :
SB6-12-7


Tot Ion: 43 Resp: 6318
Ion Ratio Lower Upper
43 100
72 24.3 22.1 33.1



#28
Bromochloromethane
Concen: 0.20 ug/l
RT: 7.38 min Scan# 1800
Delta R.T. 0.18 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58

Tgt Ion: 49 Resp: 1169
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.2
130 0.0 65.5 98.3#





#29

Tetrahydrofuran

Concen: 0.96 ug/l

RT: 7.22 min Scan# 1751

Delta R.T. 0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :

SB6-12-7





Tot Ion: 42 Resp: 1737

Ion Ratio Lower Upper

42 100

72 32.0 35.6 53.4#

71 22.3 34.0 51.0#



#30

Chloroform

Concen: 1.16 ug/l

RT: 7.37 min Scan# 1799

Delta R.T. -0.00 min

Lab File: VN049946.D




Acq: 20 Jul 2018 4:58


Tgt Ion: 83 Resp: 15102

Ion Ratio Lower Upper

83 100

85 63.2 51.4 77.2





#31

Cyclohexane

Concen: 1.01 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.02 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :

SB6-12-7


Tot Ion: 56 Resp: 10439

Ion Ratio Lower Upper

56 100

69 192.7 26.6 40.0#




84 145.4 69.0 103.6#



Abundance Ion 56.00 (55.70 to 56.70): VN049866.D (-1863) (-)

Ion 69.00 (68.70 to 69.70): VN049866.D (-1863) (-)

Ion 84.00 (83.70 to 84.70): VN049866.D (-1863) (-)



#33

1,2-Dichloroethane-d4

Concen: 52.43 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN049946.D


Acq: 20 Jul 2018 4:58

Tgt Ion: 65 Resp: 408294

Ion Ratio Lower Upper



65 100


67 52.5 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VN049866.D (-1983) (-)

Ion 66.90 (66.60 to 67.60): VN049866.D (-1983) (-)









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.59 min Scan# 2177
Delta R.T. 0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58

Instrument :
MSVOA_N
ClientSampleId :
SB6-12-7

Tot Ion:114 Resp: 955474
Ion Ratio Lower Upper
114 100
63 19.7 0.0 39.8
88 15.1 0.0 30.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VN
Ion 87.90 (87.60 to 88.60): VN






#35
Dibromofluoromethane
Concen: 44.73 ug/l
RT: 7.59 min Scan# 1867
Delta R.T. -0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58

Tgt Ion:113 Resp: 358761
Ion Ratio Lower Upper
113 100
111 103.3 81.8 122.8
192 22.3 17.8 26.6



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 3.79 ug/l
 RT: 7.67 min Scan# 1890
 Delta R.T. -0.13 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58




Instrument :
 MSVOA_N
 ClientSampleId :
 SB6-12-7


Tot Ion: 75 Resp: 40284
 Ion Ratio Lower Upper
 75 100
 110 8.1 18.1 54.4#
 77 0.0 24.9 37.3#



#38
 Carbon Tetrachloride
 Concen: 4.55 ug/l
 RT: 7.67 min Scan# 1890
 Delta R.T. -0.11 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58

Tgt Ion: 117 Resp: 51782
 Ion Ratio Lower Upper
 117 100
 119 5.8 77.2 115.8#
 121 0.0 25.0 37.6#









#43

Isopropyl Acetate
 Concen: 16.18 ug/l
 RT: 8.17 min Scan# 2046
 Delta R.T. -0.00 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58

Instrument :
 MSVOA_N
 ClientSampled :
 SB6-12-7




Tot Ion	43	Resp	206280
Ion Ratio	Lower	Upper	
43	100		
61	16.2	14.9	22.3
87	10.4	10.0	15.0




#47

Bromodichloromethane
 Concen: 0.41 ug/l
 RT: 9.40 min Scan# 2429
 Delta R.T. -0.00 min
 Lab File: VN049946.D
 Acq: 20 Jul 2018 4:58

Tgt Ion	83	Resp	4585
Ion Ratio	Lower	Upper	
83	100		
85	56.5	50.5	75.7
127	11.4	6.8	10.2#





#48

Methyl methacrylate

Concen: 1.32 ug/l

RT: 9.18 min Scan# 2361

Delta R.T. -0.02 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampled :

SB6-12-7


Tot Ion: 41 Resp: 8443

Ion Ratio Lower Upper

41 100

69 0.0 71.5 107.3#

39 18.3 44.7 67.1#




Abundance Ion 41.00 (40.70 to 41.70): VN049946.D (-2274) (-)

Ion 69.00 (68.70 to 69.70): VN049946.D (-2274) (-)

Ion 39.00 (38.70 to 39.70): VN049946.D (-2274) (-)

Time-->




Abundance Ion 41.00 (40.70 to 41.70): VN049946.D (-2274) (-)

Ion 69.00 (68.70 to 69.70): VN049946.D (-2274) (-)

Ion 39.00 (38.70 to 39.70): VN049946.D (-2274) (-)

Time-->



#50

Toluene-d8

Concen: 43.56 ug/l

RT: 10.09 min Scan# 2645

Delta R.T. 0.00 min

Lab File: VN049946.D


Acq: 20 Jul 2018 4:58

Tgt Ion: 98 Resp: 1273042

Ion Ratio Lower Upper

98 100


100 63.1 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VN049946.D (-2551) (-)

Ion 100.00 (99.70 to 100.70): VN049946.D (-2551) (-)


Time-->



Abundance Ion 98.00 (97.70 to 98.70): VN049946.D (-2551) (-)

Ion 100.00 (99.70 to 100.70): VN049946.D (-2551) (-)

Time-->



#51

4-Methyl-2-Pentanone

Concen: 0.34 ug/l

RT: 9.99 min Scan# 2612

Delta R.T. 0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :


SB6-12-7

Tot Ion: 43 Resp: 2443

Ion Ratio Lower Upper




43 100

58 38.4 31.8 47.8



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#56

Ethyl methacrylate

Concen: 2.18 ug/l

RT: 10.27 min Scan# 2700

Delta R.T. -0.16 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58


Tgt Ion: 69 Resp: 222

Ion Ratio Lower Upper

69 100

41 258.6 51.6 77.4#



39 63.5 25.8 38.8#




Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0





#59

2-Hexanone

Concen: 1.88 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :





SB6-12-7

Tot Ion: 43 Resp: 8980

Ion Ratio Lower Upper

43 100

58 24.6 27.8 83.4#



#62

4-Bromofluorobenzene

Concen: 34.34 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58




Tgt Ion: 95 Resp: 345639


Ion Ratio Lower Upper

95 100

174 91.4 0.0 179.6

176 87.3 0.0 174.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. -0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58




Instrument :
MSVOA_N
ClientSampled :
SB6-12-7


Tot Ion:117 Resp: 791336
Ion Ratio Lower Upper
117 100
82 53.9 41.9 62.9
119 32.3 25.4 38.2



#68
m/p-Xylenes
Concen: 0.23 ug/l
RT: 11.62 min Scan# 3119
Delta R.T. -0.00 min
Lab File: VN049946.D
Acq: 20 Jul 2018 4:58

Tgt Ion:106 Resp: 2826
Ion Ratio Lower Upper
106 100
91 208.0 163.4 245.2





#71

Bromoform

Concen: 0.55 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.28 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :

SB6-12-7


Tot Ion:173 Resp: 3039

Ion Ratio Lower Upper

173 100

175 614.2 24.6 73.8#




254 0.0 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. -0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58


Tgt Ion:152 Resp: 288754

Ion Ratio Lower Upper

152 100

115 56.2 28.1 84.2



150 155.9 0.0 347.6




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#76

1,2,3-Trichloropropane

Concen: 31.13 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampled :





SB6-12-7

Tot Ion: 75 Resp: 168635

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



#81

trans-1,4-Dichloro-2-butene

Concen: 97.35 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58




Tgt Ion: 75 Resp: 168635


Ion Ratio Lower Upper

75 100

53 0.2 73.9 110.9#

89 0.0 35.9 53.9#





#95

Naphthalene

Concen: 9.15 ug/l

RT: 15.14 min Scan# 4213

Delta R.T. 0.00 min

Lab File: VN049946.D

Acq: 20 Jul 2018 4:58

Instrument :

MSVOA_N

ClientSampleId :

SB6-12-7


Tot Ion:128 Resp: 69934

Ion Ratio Lower Upper

128 100

127 14.4 10.4 15.6

129 10.6 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

