

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049948.D
 Acq On : 20 Jul 2018 5:50
 Operator : MD\SY
 Sample : J4030-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB1-12-0

Quant Time: Jul 20 06:40:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	549980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	929106	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	776323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	275794	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	401486	53.11	ug/l	0.00
Spiked Amount			Recovery	=	106.22%	
35) Dibromofluoromethane	7.59	113	353067	45.27	ug/l	0.00
Spiked Amount			Recovery	=	90.54%	
50) Toluene-d8	10.09	98	1244502	43.79	ug/l	0.00
Spiked Amount			Recovery	=	87.58%	
62) 4-Bromofluorobenzene	12.40	95	335947	34.32	ug/l	0.00
Spiked Amount			Recovery	=	68.64%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.05	50	1851	0.25	ug/l	93
5) Bromomethane	2.56	94	195	Below Cal	#	19
11) Tert butyl alcohol	4.80	59	1217	2.26	ug/l	72
13) Acrolein	3.61	56	92	0.27	ug/l	1
16) Acetone	3.82	43	24602	12.57	ug/l	100
17) Carbon Disulfide	4.05	76	3628	0.20	ug/l	75
18) Methyl Acetate	4.34	43	27969	3.78	ug/l	95
20) Methylene Chloride	4.55	84	56140	7.54	ug/l	98
25) 2-Butanone	6.84	43	6376	2.28	ug/l	84
29) Tetrahydrofuran	7.21	42	962	0.55	ug/l	53
30) Chloroform	7.38	83	7390	0.58	ug/l	92
31) Cyclohexane	7.66	56	11317	1.12	ug/l	1
36) 1,1-Dichloropropene	7.67	75	39807	3.85	ug/l	47
37) Ethyl Acetate	6.84	43	6376	0.90	ug/l	68
38) Carbon Tetrachloride	7.67	117	49283	4.45	ug/l	16
41) Methacrylonitrile	7.22	41	765	0.22	ug/l	37
43) Isopropyl Acetate	8.17	43	216226	17.45	ug/l	97
47) Bromodichloromethane	9.40	83	2376	0.22	ug/l	91
48) Methyl methacrylate	9.18	41	7365	1.18	ug/l	21
51) 4-Methyl-2-Pentanone	9.99	43	2249	0.32	ug/l	86
56) Ethyl methacrylate	10.64	69	259	2.18	ug/l	1
59) 2-Hexanone	10.77	43	7400	1.59	ug/l	59
71) Bromoform	12.40	173	2572	0.48	ug/l	1
76) 1,2,3-Trichloropropane	12.40	75	165041	31.90	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.40	75	165041	99.75	ug/l	13
95) Naphthalene	15.14	128	13371	5.42	ug/l	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049948.D
Acq On : 20 Jul 2018 5:50
Operator : MD\SY
Sample : J4030-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB1-12-0

Quant Time: Jul 20 06:40:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :


SB1-12-0

Tot Ion:168 Resp: 549980

Ion Ratio Lower Upper

168 100

99 52.9 40.6 61.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

250000

200000



150000

100000

50000

0

Time-->



#3

Chloromethane

Concen: 0.25 ug/l

RT: 2.05 min Scan# 144

Delta R.T. -0.01 min

Lab File: VN049948.D


Acq: 20 Jul 2018 5:50

Tgt Ion: 50 Resp: 1851

Ion Ratio Lower Upper

50 100

52 36.8 26.1 39.1



Abundance Ion 49.90 (49.60 to 50.60): VN

Ion 51.90 (51.60 to 52.60): VN

2.05


1500


1000

500

0

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 2.56 min Scan# 303

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :


SB1-12-0

Tot Ion: 94 Resp: 195

Ion Ratio Lower Upper




94 100

96 16.2 75.7 113.5#



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0



#11

Tert butyl alcohol

Concen: 2.26 ug/l

RT: 4.80 min Scan# 999

Delta R.T. 0.00 min

Lab File: VN049948.D


Acq: 20 Jul 2018 5:50

Tgt Ion: 59 Resp: 1217

Ion Ratio Lower Upper



59 100


57 0.0 8.5 12.7#



Abundance Ion 59.00 (58.70 to 59.70): VN0

Ion 57.00 (56.70 to 57.70): VN0





#13

Acrolein

Concen: 0.27 ug/l

RT: 3.61 min Scan# 627

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampled :





SB1-12-0

Tot Ion: 56 Resp: 92

Ion Ratio Lower Upper

56 100

55 259.8 55.8 83.8#



#16

Acetone

Concen: 12.57 ug/l

RT: 3.82 min Scan# 695

Delta R.T. 0.00 min

Lab File: VN049948.D




Acq: 20 Jul 2018 5:50


Tgt Ion: 43 Resp: 24602

Ion Ratio Lower Upper

43 100

58 32.3 25.9 38.9









#17
Carbon Disulfide
Concen: 0.20 ug/l
RT: 4.05 min Scan# 764
Delta R.T. 0.00 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50




Instrument :
MSVOA_N
ClientSampled :
SB1-12-0


Tot Ion: 76 Resp: 3628
Ion Ratio Lower Upper
76 100
78 0.0 7.1 10.7#



#18
Methyl Acetate
Concen: 3.78 ug/l
RT: 4.34 min Scan# 854
Delta R.T. 0.01 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion: 43 Resp: 27969
Ion Ratio Lower Upper
43 100
74 22.5 19.8 29.8









#20
Methvlene Chloride
Concen: 7.54 ug/l
RT: 4.55 min Scan# 922
Delta R.T. 0.01 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Instrument :
MSVOA_N
ClientSampleId :
SB1-12-0

Tot Ion:	84	Resp:	56140
Ion	Ratio	Lower	Upper
84	100		
49	119.8	95.4	143.2
51	39.8	28.4	42.6
86	64.8	52.6	79.0




Abundance Ion 83.90 (83.60 to 84.60): VN0






#25
2-Butanone
Concen: 2.28 ug/l
RT: 6.84 min Scan# 1633
Delta R.T. 0.00 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion:	43	Resp:	6376
Ion	Ratio	Lower	Upper
43	100		
72	19.4	22.1	33.1#



Abundance Ion 43.00 (42.70 to 43.70): VN0









#29
Tetrahydrofuran
Concen: 0.55 ug/l
RT: 7.21 min Scan# 1748
Delta R.T. -0.01 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50




Instrument :
MSVOA_N
ClientSampled :
SB1-12-0


Tot Ion: 42 Resp: 962
Ion Ratio Lower Upper
42 100
72 3.0 35.6 53.4#
71 24.0 34.0 51.0#



#30
Chloroform
Concen: 0.58 ug/l
RT: 7.38 min Scan# 1800
Delta R.T. 0.00 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion: 83 Resp: 7390
Ion Ratio Lower Upper
83 100
85 70.8 51.4 77.2





#31

Cyclohexane

Concen: 1.12 ug/l

RT: 7.66 min Scan# 1888

Delta R.T. 0.01 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :

SB1-12-0


Tot Ion: 56 Resp: 11317

Ion Ratio Lower Upper

56 100

69 175.5 26.6 40.0#




84 128.4 69.0 103.6#



Abundance Ion 56.00 (55.70 to 56.70): VN049866.D (-1863) (-)

Ion 69.00 (68.70 to 69.70): VN049866.D (-1863) (-)

Ion 84.00 (83.70 to 84.70): VN049866.D (-1863) (-)



#33

1,2-Dichloroethane-d4

Concen: 53.11 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN049948.D


Acq: 20 Jul 2018 5:50

Tgt Ion: 65 Resp: 401486

Ion Ratio Lower Upper



65 100


67 53.0 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VN049866.D (-1983) (-)

Ion 66.90 (66.60 to 67.60): VN049866.D (-1983) (-)





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 8.59 min Scan# 2176

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampled :

SB1-12-0


Tot Ion:114 Resp: 929106

Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.8




88 15.3 0.0 30.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0



#35

Dibromofluoromethane

Concen: 45.27 ug/l

RT: 7.59 min Scan# 1867

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50


Tgt Ion:113 Resp: 353067

Ion Ratio Lower Upper

113 100

111 103.0 81.8 122.8



192 22.4 17.8 26.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 3.85 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. -0.13 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampled :

SB1-12-0





Tot Ion: 75 Resp: 39807

Ion Ratio Lower Upper

75 100

110 7.2 18.1 54.4#

77 0.0 24.9 37.3#



#37

Ethyl Acetate

Concen: 0.90 ug/l

RT: 6.84 min Scan# 1633

Delta R.T. -0.09 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50




Tgt Ion: 43 Resp: 6376


Ion Ratio Lower Upper

43 100

61 0.0 11.7 17.5#

70 0.0 8.4 12.6#









#38
Carbon Tetrachloride
Concen: 4.45 ug/l
RT: 7.67 min Scan# 1891
Delta R.T. -0.10 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50




Instrument :
MSVOA_N
ClientSampled :
SB1-12-0


Tot Ion:117 Resp: 49283
Ion Ratio Lower Upper
117 100
119 6.1 77.2 115.8#
121 0.0 25.0 37.6#



#41
Methacrylonitrile
Concen: 0.22 ug/l
RT: 7.22 min Scan# 1752
Delta R.T. 0.05 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion: 41 Resp: 765
Ion Ratio Lower Upper
41 100
39 38.3 45.4 68.2#
67 0.0 65.4 98.0#
52 0.0 24.7 37.1#





#43

Isopropyl Acetate

Concen: 17.45 ug/l

RT: 8.17 min Scan# 2046

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :

SB1-12-0


Tot Ion: 43 Resp: 216226

Ion Ratio Lower Upper

43 100

61 17.1 14.9 22.3




87 11.1 10.0 15.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#47

Bromodichloromethane

Concen: 0.22 ug/l

RT: 9.40 min Scan# 2430

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50


Tgt Ion: 83 Resp: 2376

Ion Ratio Lower Upper

83 100

85 70.0 50.5 75.7



127 13.5 6.8 10.2#




Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 84.90 (84.60 to 85.60): VN0

Ion 126.80 (126.50 to 127.50): V









#48
Methyl methacrylate
Concen: 1.18 ug/l
RT: 9.18 min Scan# 2362
Delta R.T. -0.02 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50




Instrument :
MSVOA_N
ClientSampled :
SB1-12-0


Tot Ion: 41 Resp: 7365
Ion Ratio Lower Upper
41 100
69 0.0 71.5 107.3#
39 16.4 44.7 67.1#



#50
Toluene-d8
Concen: 43.79 ug/l
RT: 10.09 min Scan# 2644
Delta R.T. 0.00 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion: 98 Resp: 1244502
Ion Ratio Lower Upper
98 100
100 63.4 51.7 77.5





#51

4-Methyl-2-Pentanone

Concen: 0.32 ug/l

RT: 9.99 min Scan# 2612

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampled :





SB1-12-0

Tot Ion: 43 Resp: 2249

Ion Ratio Lower Upper

43 100

58 31.3 31.8 47.8#



#56

Ethyl methacrylate

Concen: 2.18 ug/l

RT: 10.64 min Scan# 2816

Delta R.T. 0.21 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50




Tgt Ion: 69 Resp: 259


Ion Ratio Lower Upper

69 100

41 923.2 51.6 77.4#

39 512.4 25.8 38.8#





#59

2-Hexanone

Concen: 1.59 ug/l

RT: 10.77 min Scan# 2854

Delta R.T. 0.01 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampled :





SB1-12-0

Tot Ion: 43 Resp: 7400

Ion Ratio Lower Upper

43 100

58 25.9 27.8 83.4#



#62

4-Bromofluorobenzene

Concen: 34.32 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50




Tgt Ion: 95 Resp: 335947


Ion Ratio Lower Upper

95 100

174 91.6 0.0 179.6

176 87.4 0.0 174.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50




Instrument :
MSVOA_N
ClientSampled :
SB1-12-0


Tot Ion:117 Resp: 776323
Ion Ratio Lower Upper
117 100
82 53.4 41.9 62.9
119 32.1 25.4 38.2



#71
Bromoform
Concen: 0.48 ug/l
RT: 12.40 min Scan# 3362
Delta R.T. 0.27 min
Lab File: VN049948.D
Acq: 20 Jul 2018 5:50

Tgt Ion:173 Resp: 2572
Ion Ratio Lower Upper
173 100
175 748.9 24.6 73.8#
254 0.0 0.2 0.2#





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :

SB1-12-0


Tot Ion:152 Resp: 275794

Ion Ratio Lower Upper

152 100

115 56.8 28.1 84.2




150 157.7 0.0 347.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 31.90 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN049948.D


Acq: 20 Jul 2018 5:50

Tgt Ion: 75 Resp: 165041

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 99.75 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50

Instrument :

MSVOA_N

ClientSampleId :

SB1-12-0


Tot Ion: 75 Resp: 165041

Ion Ratio Lower Upper

75 100

53 0.2 73.9 110.9#




89 0.0 35.9 53.9#



Abundance Ion 74.90 (74.60 to 75.60): VN049948.D (-3363) (-)

Ion 53.00 (52.70 to 53.70): VN049948.D (-3363) (-)

Ion 88.90 (88.60 to 89.60): VN049948.D (-3363) (-)



#95

Naphthalene

Concen: 5.42 ug/l

RT: 15.14 min Scan# 4213

Delta R.T. 0.00 min

Lab File: VN049948.D

Acq: 20 Jul 2018 5:50


Tgt Ion: 128 Resp: 13371

Ion Ratio Lower Upper

128 100

127 13.6 10.4 15.6

129 5.2 8.6 13.0#



Abundance Ion 128.00 (127.70 to 128.70): VN049948.D (-4213) (-)

Ion 127.00 (126.70 to 127.70): VN049948.D (-4213) (-)

Ion 129.00 (128.70 to 129.70): VN049948.D (-4213) (-)

