

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049918.D
 Acq On : 19 Jul 2018 15:58
 Operator : MD\SY
 Sample : PB111251ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111251ZHE#05

Quant Time: Jul 20 02:07:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	577507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	954383	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	814654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	302312	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	410115	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
35) Dibromofluoromethane	7.59	113	363509	45.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.74%	
50) Toluene-d8	10.09	98	1288783	44.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.30%	
62) 4-Bromofluorobenzene	12.40	95	353420	35.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.30%	
Target Compounds						
16) Acetone	3.82	43	17141	8.34	ug/l	94
18) Methyl Acetate	4.34	43	18899	2.14	ug/l	97
20) Methylene Chloride	4.55	84	38983	4.99	ug/l	97
43) Isopropyl Acetate	8.17	43	298284	23.43	ug/l	94
59) 2-Hexanone	10.77	43	20341	4.26	ug/l #	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049918.D
 Acq On : 19 Jul 2018 15:58
 Operator : MD\SY
 Sample : PB111251ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111251ZHE#05

Quant Time: Jul 20 02:07:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

