

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049925.D
 Acq On : 19 Jul 2018 19:00
 Operator : MD\SY
 Sample : PB111251ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111251ZHE#12

Quant Time: Jul 20 02:21:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	548261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	941945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	787433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	289513	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	404280	53.65	ug/l	0.00
Spiked Amount						
			Recovery	=	107.30%	
35) Dibromofluoromethane	7.59	113	356981	45.14	ug/l	0.00
Spiked Amount						
			Recovery	=	90.28%	
50) Toluene-d8	10.09	98	1257192	43.63	ug/l	0.00
Spiked Amount						
			Recovery	=	87.26%	
62) 4-Bromofluorobenzene	12.40	95	341313	34.40	ug/l	0.00
Spiked Amount						
			Recovery	=	68.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	16632	8.52	ug/l	97
18) Methyl Acetate	4.34	43	29504	4.05	ug/l	99
20) Methylene Chloride	4.55	84	43160	5.82	ug/l	98
43) Isopropyl Acetate	8.17	43	364282	28.99	ug/l #	89
59) 2-Hexanone	10.77	43	21233	4.50	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049925.D
Acq On : 19 Jul 2018 19:00
Operator : MD\SY
Sample : PB111251ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111251ZHE#12

Quant Time: Jul 20 02:21:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

