

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049947.D
 Acq On : 20 Jul 2018 5:24
 Operator : MD\SY
 Sample : J4030-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB7-12-7

Quant Time: Jul 20 06:46:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	536428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	902324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	750128	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	264984	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	390832	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
35) Dibromofluoromethane	7.59	113	345351	45.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.18%	
50) Toluene-d8	10.09	98	1208443	43.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.56%	
62) 4-Bromofluorobenzene	12.40	95	324908	34.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.36%	
Target Compounds						
16) Acetone	3.82	43	22803	11.95	ug/l	98
18) Methyl Acetate	4.34	43	27456	3.81	ug/l	92
20) Methylene Chloride	4.55	84	67077	9.24	ug/l	98
30) Chloroform	7.37	83	15505	1.26	ug/l	96
43) Isopropyl Acetate	8.17	43	103554	8.60	ug/l #	92
95) Naphthalene	15.13	128	7976	5.07	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049947.D
Acq On : 20 Jul 2018 5:24
Operator : MD\SY
Sample : J4030-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB7-12-7

Quant Time: Jul 20 06:46:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

