

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049953.D
 Acq On : 20 Jul 2018 7:59
 Operator : MD\SY
 Sample : J4016-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB4-U-23-13

Quant Time: Jul 20 08:24:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	571481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	947945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	797055	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	285064	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	411512	52.39	ug/l	0.00
Spiked Amount			Recovery	=	104.78%	
35) Dibromofluoromethane	7.59	113	361107	45.38	ug/l	0.00
Spiked Amount			Recovery	=	90.76%	
50) Toluene-d8	10.09	98	1287046	44.39	ug/l	0.00
Spiked Amount			Recovery	=	88.78%	
62) 4-Bromofluorobenzene	12.40	95	349914	35.04	ug/l	0.00
Spiked Amount			Recovery	=	70.08%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	21483	10.56	ug/l	96
18) Methyl Acetate	4.34	43	23365	2.88	ug/l	98
20) Methylene Chloride	4.55	84	53748	6.95	ug/l	99
43) Isopropyl Acetate	8.17	43	221961	17.55	ug/l	97
59) 2-Hexanone	10.77	43	15846	3.34	ug/l #	61
95) Naphthalene	15.14	128	64063	8.81	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049953.D
Acq On : 20 Jul 2018 7:59
Operator : MD\SY
Sample : J4016-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB4-U-23-13

Quant Time: Jul 20 08:24:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

