

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049905.D
 Acq On : 19 Jul 2018 9:42
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 20 01:35:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.534	0.458	14.2	103	0.00
3 P	Chloromethane	0.682	0.567	16.9	94	0.00
4 C	Vinyl Chloride	0.728	0.638	12.4#	97	0.00
5 T	Bromomethane	0.471	0.407	13.6	101	0.00
6 T	Chloroethane	0.475	0.413	13.1	98	0.00
7 T	Trichlorofluoromethane	1.001	0.890	11.1	99	0.00
8 T	Diethyl Ether	0.349	0.310	11.2	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.565	8.4	104	0.00
10 T	Methyl Iodide	0.828	0.785	5.2	97	0.00
11 T	Tert butyl alcohol	0.049	0.039	20.4#	81	0.00
12 CM	1,1-Dichloroethene	0.542	0.491	9.4#	99	0.00
13 T	Acrolein	0.032	0.025	21.9#	73	0.00
14 T	Allyl chloride	0.855	0.804	6.0	99	0.00
15 T	Acrylonitrile	0.215	0.188	12.6	88	0.00
16 T	Acetone	0.178	0.193	-8.4	123	0.00
17 T	Carbon Disulfide	1.622	1.400	13.7	97	0.00
18 T	Methyl Acetate	0.672	0.510	24.1#	93	0.00
19 T	Methyl tert-butyl Ether	1.492	1.435	3.8	96	0.00
20 T	Methylene Chloride	0.677	0.580	14.3	98	0.00
21 T	trans-1,2-Dichloroethene	0.576	0.541	6.1	98	0.00
22 T	Diisopropyl ether	1.744	1.722	1.3	97	0.00
23 T	Vinyl Acetate	1.224	1.184	3.3	93	0.00
24 P	1,1-Dichloroethane	1.119	1.012	9.6	98	0.00
25 T	2-Butanone	0.254	0.258	-1.6	102	0.00
26 T	2,2-Dichloropropane	0.871	0.887	-1.8	111	0.00
27 T	cis-1,2-Dichloroethene	0.632	0.618	2.2	99	0.00
28 T	Bromochloromethane	0.506	0.482	4.7	96	0.00
29 T	Tetrahydrofuran	0.160	0.144	10.0	85	0.00
30 C	Chloroform	1.149	1.067	7.1#	100	0.00
31 T	Cyclohexane	1.047	0.876	16.3	101	0.00
32 T	1,1,1-Trichloroethane	0.993	0.905	8.9	99	0.00
33 S	1,2-Dichloroethane-d4	0.687	0.654	4.8	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.420	0.416	1.0	99	0.00
36 T	1,1-Dichloropropene	0.556	0.562	-1.1	99	0.00
37 T	Ethyl Acetate	0.382	0.346	9.4	87	0.00
38 T	Carbon Tetrachloride	0.596	0.575	3.5	99	0.00
39 T	Methylcyclohexane	0.574	0.615	-7.1	104	0.00
40 TM	Benzene	1.694	1.680	0.8	99	0.00
41 T	Methacrylonitrile	0.191	0.189	1.0	98	0.00
42 TM	1,2-Dichloroethane	0.561	0.546	2.7	98	0.00
43 T	Isopropyl Acetate	0.667	0.638	4.3	91	0.00
44 TM	Trichloroethene	0.444	0.442	0.5	101	0.00
45 C	1,2-Dichloropropane	0.457	0.443	3.1#	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
 Data File : VN049905.D
 Acq On : 19 Jul 2018 9:42
 Operator : MD\SY
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 20 01:35:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 01:58:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.283	0.271	4.2	96	0.00
47 T	Bromodichloromethane	0.586	0.579	1.2	98	0.00
48 T	Methyl methacrylate	0.335	0.319	4.8	90	0.00
49 T	1,4-Dioxane	0.005	0.004	20.0	82	0.00
50 S	Toluene-d8	1.529	1.553	-1.6	98	0.00
51 T	4-Methyl-2-Pentanone	0.373	0.359	3.8	87	0.00
52 CM	Toluene	1.011	1.058	-4.6#	99	0.00
53 T	t-1,3-Dichloropropene	0.580	0.599	-3.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.641	0.675	-5.3	100	0.00
55 T	1,1,2-Trichloroethane	0.418	0.391	6.5	96	0.00
56 T	Ethyl methacrylate	0.487	0.501	-2.9	92	0.00
57 T	1,3-Dichloropropane	0.658	0.649	1.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.222	0.243	-9.5	93	0.00
59 T	2-Hexanone	0.250	0.282	-12.8	99	0.00
60 T	Dibromochloromethane	0.458	0.452	1.3	97	0.00
61 T	1,2-Dibromoethane	0.390	0.378	3.1	93	0.00
62 S	4-Bromofluorobenzene	0.527	0.548	-4.0	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.449	0.439	2.2	101	0.00
65 PM	Chlorobenzene	1.256	1.243	1.0	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.487	0.471	3.3	99	0.00
67 C	Ethyl Benzene	2.038	2.159	-5.9#	99	0.00
68 T	m/p-Xylenes	0.771	0.836	-8.4	99	0.00
69 T	o-Xylene	0.741	0.782	-5.5	97	0.00
70 T	Styrene	1.200	1.334	-11.2	98	0.00
71 P	Bromoform	0.348	0.337	3.2	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.798	3.903	-2.8	99	0.00
74 T	N-amyl acetate	1.156	1.084	6.2	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.138	0.962	15.5	90	0.00
76 T	1,2,3-Trichloropropane	0.938	0.774	17.5	84	0.00
77 T	Bromobenzene	1.049	1.002	4.5	98	0.00
78 T	n-propylbenzene	4.318	4.574	-5.9	100	0.00
79 T	2-Chlorotoluene	2.651	2.687	-1.4	99	0.00
80 T	1,3,5-Trimethylbenzene	3.044	3.235	-6.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.279	7.0	90	0.00
82 T	4-Chlorotoluene	2.660	2.744	-3.2	99	0.00
83 T	tert-Butylbenzene	2.673	2.758	-3.2	97	0.00
84 T	1,2,4-Trimethylbenzene	3.084	3.330	-8.0	99	0.00
85 T	sec-Butylbenzene	3.533	3.830	-8.4	101	0.00
86 T	p-Isopropyltoluene	3.012	3.321	-10.3	100	0.00
87 T	1,3-Dichlorobenzene	1.827	1.802	1.4	99	0.00
88 T	1,4-Dichlorobenzene	1.832	1.764	3.7	99	0.00
89 T	n-Butylbenzene	2.487	2.813	-13.1	103	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071918\
Data File : VN049905.D
Acq On : 19 Jul 2018 9:42
Operator : MD\SY
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jul 20 01:35:01 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071818W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 01:58:39 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.676	0.616	8.9	100	0.00
91 T	1,2-Dichlorobenzene	1.797	1.739	3.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.176	0.146	17.0	88	0.00
93 T	1,2,4-Trichlorobenzene	0.825	0.927	-12.4	99	0.00
94 T	Hexachlorobutadiene	0.589	0.575	2.4	107	0.00
95 T	Naphthalene	1.828	1.904	-4.2	86	0.00
96 T	1,2,3-Trichlorobenzene	0.862	0.899	-4.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6